

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130579</u>	
Name	<u>C. Ritala</u>	ID#	<u>3791</u> Date <u>5-27-08</u>
A Agency	<u>APD</u>	Phone #	<u>786-2640</u>
Instrument Location <u>Anchorage Jail</u>			
Alco S/N <u>57447</u>		Target Value	<u>077</u> High Pressure <u>850</u>
<u>.0808</u> ^{6/12/05}		<u>.080</u> ^{8/6/12/08}	
Alco Test Value Average <u>80</u>		<u>80</u>	
1 st Alco		2 nd Alco	
Signature <u>[Signature]</u> /1249			
(OVER)			

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[Signature] 6/12/08
Date
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska.

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record

Serial Number 130579 ✓ Model DataMaster
Location ANC1 ✓ Anchorage Police Dept. ✓

Diagnostic Test Record

Serial Number 130579 ✓ Model DataMaster
Location ANC1 ✓ Anchorage Police Dept.

Subject Information

Last Name Verification Of Calibration O.L. # A
First Name A

Diagnostic Test Results

Test Date 05/27/2008 Computer Okay
Test Time 22:39:28 Program Okay
Software Date 02/20/2001
Sample Chamber (C) 50
Breath Tube (C) 41
Barometer (in) 30.30
Flow Detector Okay
High Speed Pump Okay
Detector Okay
Filter Okay
Quartz Okay
Calibration Okay

Operator Information

Operator Name Ritala Case A
Operator # 3791 Department ANC1

Test Results

Alco Target .077 Alco Serial 57447
Nominal 0.078 Barometer 30.32 Test Type V
Dry Gas 1 0.080 ✓ Time: 22:44:31 Test Date 05/27/2008
Subject Alcohol 0.000 Time: 22:45:47 Test Time 22:42:53
Dry Gas 2 0.080 ✓ Time: 22:47:08 Report Y