

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130579</u> ✓
Name	<u>Colleen O'Bryant</u>	ID#	<u>6237</u> ✓
		Date	<u>8/20/10</u> ✓
A	Agency	<u>SCDL</u>	Phone # <u>269-5783</u>
	Instrument Location	<u>SCDL 8/20/10 SCDL</u>	
B	Alco S/N	<u>500408</u> ✓	Target Value <u>0.080</u> ✓
			High Pressure <u>300psi</u>
	Alco Test Values	<u>0.079</u> ✓ 1st Alco	<u>0.079</u> ✓ 2nd Alco
	Signature	<u>C. O'Bryant</u>	

BMB
8/27/10

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 5th day of Oct, 2010.

Daniel L. Nicholas (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

AUGUST 20, 2010

OPERATOR'S NAME:
O'BRYAN/COLLEEN/S
OPERATOR'S NUMBER: 6237
SUBJECT'S LAST NAME:

VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 023456
DEPT/AGENCY: AAA1
CASE/REPORT: 10-1234
TEST TYPE: V
ALCO TARGET VALUE: .080
ALCO S/N: SCOTT00

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.81 in
ALCO TARGET .079 08:47
BLANK TEST .000 08:48
INTERNAL STANDARD VERIFIED 08:48
ALCO TV 29.81 in .079 08:49
BLANK TEST .000 08:49
SUBJECT SAMPLE .000 08:50
BLANK TEST .000 08:50
ALCO TV 29.81 in .079 08:51
BLANK TEST .000 08:51

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

AUGUST 20, 2010
TIME 08:44

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/00
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 42c
BAROMETER: 29.81 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~