

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130584

Supervisor/Operator Performing the Verification Procedure:

Name Supt. James R. R. Rea

ID# 3161

Date 06-24-08 ✓

A Agency Kotzebue Police Department ✓

Phone # 907-442-3352

Instrument Location Kotzebue Regional Jail 258d rm#107 Fourth Avenue Kotzebue, AK 99752

Alco S/N 78890 ✓

Target Value .082 ✓

High Pressure 550


B

Alco Test Values:

.082 ✓
1st Alco

.083 ✓
2nd Alco

Signature _____


Supt. James R. R. Rea
(OVER)


7/7/08

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Rea, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

7/8/08
Date

BT9 03/09
JUL 03 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130584

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

JUNE 24, 2000 ✓ *COS #77/08*

OPERATOR'S NAME:
REA, JAMES RR

OPERATOR'S NUMBER: 3161

SUBJECT'S LAST NAME:
VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :
VERIFICATION OF CAL

G.L. #: 06242000

DEPT/AGENCY: 0T21

CASE/REPORT: 06242000

TEST TYPE: V

ALCO TARGET VALUE: 002 ✓

ALCO S/N: 78890 ✓

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.96 in

ALCO TARGET	.001	18:44
BLANK TEST	.000	18:45
INTERNAL STANDARD	VERIFIED	18:45
ALCO TV 29.96 in	.002 ✓	18:46
BLANK TEST	.000 ✓	18:46
SUBJECT SAMPLE	.000 ✓	18:49
BLANK TEST	.000 ✓	18:49
ALCO TV 29.96 in	.002 ✓	18:50
BLANK TEST	.000	18:51

20/21

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

JUNE 24, 2000 ✓

TIME 18:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.96 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""#%&'()*+,-./0123456789:;<=>?@ABCDEFB
@HILKLMNOPQRSTUWVXYZ[\]^_`abc.de fghi.jklmno
{pqr.stuvwxyz{|}~"