

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

AUG 05 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130584

Supervisor/Operator Performing the Verification Procedure:

Name Supt. James R. R. Rea

ID# 3161

Date 07-26-09

A Agency Kotzebue Police Department

Phone # 907-442-3352

Instrument Location Kotzebue Regional Jail 258d rm#107 Fourth Avenue Kotzebue, AK 99752

Alco S/N 78890

Target Value .082

High Pressure 1500

Alco Test Values:

B

1st Alco

.082 - .081
CMW
8/7/09

2nd Alco .081

Signature _____

[Signature]
Supt. James R. R. Rea
(OVER)

CMW
8/7/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of August, 2009.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130584

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

JULY 06, 2009
TIME 08:12

OPERATOR'S NUMBER: 3161
SUBJECT'S LAST NAME: _____
VERIFICATION OF CAL _____
SUBJECT'S FIRST NAME/MI: _____
VERIFICATION OF CAL _____
O.L. #: 07262009
DEPT/AGENCY: OTZ1
CASE/REPORT: 07262009
TEST TYPE: V
ALCO TARGET VALUE: 002
ALCO S/N: 78890

CSB
8/17/09

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.03 in		
ALCO TARGET	.002	08:05
BLANK TEST	.000	08:06
INTERNAL STANDARD	VERIFIED	08:07
ALCO TV 30.03 in	.001	08:07
BLANK TEST	.000	08:08
SUBJECT SAMPLE	.000	08:09
BLANK TEST	.000	08:09
ALCO TV 30.04 in	.001	08:10
BLANK TEST	.000	08:11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

JULY 06, 2009
TIME 08:12

DIAGNOSTIC CHECK

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS: _____
SAMPLE CHAMBER: 49
BREATH TUBE: 41C

BAROMETER: 30.04 in

FLOW DETECTOR: OKAY

PUMP HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () * +