

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130666
Name		Gary Yamamoto	ID# 2765 Date 6-20-08 ✓
A	Agency	Fairbanks Police Department ✓	Phone # 459-6500
Instrument Location 911 Cushman Street Fairbanks, Alaska 99701			
Alco S/N		X301737 ✓	Target Value .080 ✓ High Pressure 800
B	Alco Test Value Average	.084 ✓	.084 ✓
		1 st Alco	2 nd Alco
Signature			CSB 7/3/08


(OVER)

Under the Alaska Rules of Evidence, I certify that;
(Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Gary Yamamoto, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

7/7/08

Date

JUN 26 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N _____ 130666 _____

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

JUNE 20, 2008

OPERATOR'S NAME:

YAMAMOTO

OPERATOR'S NUMBER: 2765

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 11111

DEPT/AGENCY: FBK1

CASE/REPORT: 11111

TEST TYPE: V

ALCO TARGET VALUE: .000 ✓

ALCO S/N: X301737 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.75 in
ALCO TARGET .079 16:44
BLANK TEST .000 16:45
INTERNAL STANDARD VERIFIED 16:45
ALCO TV 29.75 in .004 ✓ 16:46
BLANK TEST .000 16:46
SUBJECT SAMPLE .000 16:47
BLANK TEST .000 16:47
ALCO TV 29.75 in .004 ✓ 16:48
BLANK TEST .000 16:49

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

JUNE 20, 2008 ✓

TIME 16:42

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 29.75 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~?