

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Effective: 11/10/2021

Version: 2.0

Date: 11/19/2021 Instrument #: 100360

Instrument Location: Sand Point AK. Booking Room

Supervisor Name and Agency: Richard Lowery Sand Point Police Dept.

Describe the Status Change or Issue with the Instrument:

Internal Standard error; Followed by Fatal System Error; followed by a black screen with 5:00 min. reset clock.

Removed from service by replacement instrument 100384

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Unk

Out of Service Date/Time: 11/19/2021 @ 1030 hrs

**For Use by SCDL**

Additional Notes

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740