

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/12/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.98: 0.077
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 131 psi

BLANK TEST 0.000 12:02
INTERNAL STANDARD VERIFIED 12:02

Gas Flow Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO

Flow Rate = 4.478 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

File Only dm 01/02/25