

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 5/18/17 Instrument Serial Number: 100390  
~~03/03/17~~

Instrument Location: AST/Emmonak

Supervisor Name and Agency: Colleen O'Bryant /SCDL

Describe the Status Change or Issue with the Instrument:

External Standard expired  
01315080A2  
Exp 2/5/17

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 5/18/17

Out of Service Date/Time: 2/5/17

For Use by SCDL

Additional Notes

Tank was expired (starting 2/5/17, new tank)  
uninstalled 5/18/17 and VQE passed. No subject  
tests performed between 2/5/17 and 5/18/17. 03/10/17

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740