

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 4/11/2023

Instrument #: 100361

Instrument Location: Nome PD

Supervisor Name and Agency: Wade Harrison, Nome PD

Describe the Status Change or Issue with the Instrument:

Out of Service, sent back to Crime lab

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 3-31-23 approx noon.

For Use by SCDL

Additional Notes

Instrument was pulled out of service due to multiple fatal error and interference detected status messages. The instrument also had low external standard values on the 02-22-2023 VOC. TLM 04/11/2023

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor