

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 8-4-18 Instrument Serial Number: 100410

Instrument Location: Saint Marys

Supervisor Name and Agency: Trooper Melvin Colley

Describe the Status Change or Issue with the Instrument:

Taken out of service to be sent to crime lab

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 8-4-18 around 1500 hours

For Use by SCDL

Additional Notes

Previous tank leaked out. Potentially repaired in field by o-ring replacement.
Seems to be holding pressure via review of XML files.
JMT 08/07/18

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740