

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMCIF 2015 R1
Status: Active

Instrument S/N 100377

Received From Crime Lab/Salmon

Date 12/22/16

Analyst Brandi Barnett

1. Reason for instrument return:

Removed from service for persistent gas flow errors. This was an inhouse instrument.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax N/A

4. Upload N/A

5. Update agency information and polling status in DM Host Calling Program N/A

6. Check for VOCs N/A

From NPAS

7. NPAS Repair Paperwork (attached)