

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 10/25/21 Instrument Serial Number: 100384

Instrument Location: Crime Lab (Salmon)

Supervisor Name and Agency: Derek Watten - SCDL

Describe the Status Change or Issue with the Instrument:

Checked instrument for VOC day and found an internal standard error status message. Attempted to run VOC again and received another internal standard error. Removed instrument from use.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100384

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/24/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.18: 0.078
LOT #: AG018102
EXPIRATION: 06/29/2022
TANK PRESSURE: 383 psi

BLANK TEST 0.000 12:02
INTERNAL STANDARD ERROR 12:02

Internal Standard Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED
Breath Tube = 47.1°C PASSED

PUMP INFO

Flow Rate = 4.724 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED