

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 10/10/2025

Instrument #: 100692

Instrument Location: ANC Airport Police Department

Supervisor Name and Agency: Cody Fenton, ANC Airport Police

Describe the Status Change or Issue with the Instrument:

Fatal error status message occurred on 10/10/2025. At approximately 2017 hours, I began conducting a "Test" on the DataMaster DMT to find no error message and working without issue.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: ~~10/10/2025~~ COB 2025.10.30

Out of Service Date/Time: Click or tap here to enter text.

For Use by SCDL

Additional Notes

Click or tap here to enter text.

I spoke with Officer Cody Fenton regarding the status message of Fatal Error and informed him it is a known software issue for the instrument to restart when the blank space for test type is selected.

100692 was not taken out of service and remains in service.

COB 2025.10.30

Email completed form to dps.scdl.toxicology@alaska.gov.