

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100364

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/07/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.71: 0.079
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 1213 psi

BLANK TEST	0.000	15:38
INTERNAL STANDARD	VERIFIED	15:38
EXTERNAL STANDARD	0.080	15:39
BLANK TEST	0.000	15:40
EXTERNAL STANDARD	0.081	15:40
BLANK TEST	0.000	15:41
EXTERNAL STANDARD	0.080	15:41
BLANK TEST	0.000	15:42
EXTERNAL STANDARD	0.081	15:42
BLANK TEST	0.000	15:43
EXTERNAL STANDARD	0.080	15:43
BLANK TEST	0.000	15:44

Average = 0.0804
Std Dev = 0.0005

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 47.6°C	PASSED

PUMP INFO

Flow Rate = 4.570 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO


Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 10/25/19

Charles R. Foster
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 25 day of 10, 2019


Notary Public
My Commission Expires With Office



JM 09/10/19