

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMCIF 2015 R1  
Status: Active

Instrument S/N 100364

Received From NPAS

Date 10/23/15

Analyst Brandi Barnett

1. Reason for instrument return:

Return from NPAS

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax  N/A

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs  N/A

From NPAS

7. NPAS Repair Paperwork (attached)

Work Ord No 49975  
Date 9/23/15

National Patent Analytical Systems, Inc.  
Service Action Report

Serial No: 100364

Customer: Alaska Dept of Public

Phone: 907-269-5582

Contact: Bolz

War Exp: 1/13/13

Mx Exp:

Ship Damage?  Y  N

Cust Pick Up??  Y  N

Initial Comments: note any shipping damage or estimates required, along with accessories

Describe problem seen and/or attach documents received

see attached document

Keyword:

Finding:

Error Code:

Enter Hours:

Repair Comments and work Performed

replaced chopper wheel and chopper motor, calibrated and preformed several accuracy tests with no problems.

Solution Lot No:

Voltage Set Values In		Voltage Set Values Out	
Lamp	Brth Vol	Lamp	Brth Vol
Meter	Filter Dif	Meter	Filter Dif
Stability		Stability	

Part No	Qty Ord	Pick	Back Ordered	Total ship	Description	Price Ea
31250	1.0				WHEEL, CHOPPER 8 WINDOW, BLK OXIDE warranty	0.000
41310	1.0				MOTOR ASS'Y, 12X20.6MM warranty	0.000
9DMT	1.0				DATAMASTER DMT SERVICE RETURN 100364	
NOTE	1.0				NANUK CASE	
r500-1	2.0				Service Charge for Repair warranty	0.000

Date Completed: 10/12/15

Technician: M. BARBELY

Pick up Signature: