

	<b>State of Alaska Scientific Crime Detection Laboratory</b>		
	<b>DataMaster Acceptance Form -- DMAF2011R0</b>		
	Issuing Authority -- Nita Boiz, Scientific Director		
Approval date: 03/01/11	Effective date: 03/01/11	Archive date: Active	

DataMaster DMT Acceptance Form

Instrument # 100574

Date: 5/10/11

Analyst: C. O. Bryant

1. Check in instrument in to JusticeTrax Breath Alcohol Database
2. Visual Inspection
3. Update software
4. Print calibration factors (attached)
5. NPAS calibration paperwork (attached)
6. Barcode Instrument
7. State of Alaska Property ID# 10140883

**B-100374** 100374



GSNNV37RSHARM B-100374  
Breath Program

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dnt: 100374  
-----

Date: 03/01/2011  
Time: 14:10:25

Ca	=	0.1000			
CAL	=	0.982183	0.800	<=	CAL < 1.200
b1	=	0.0006	0.0000	<=	b1 < 0.0040
b2	=	0.0058	0.0010	<=	b2 < 0.0100
b3	=	0.0001	0.0000	<=	b3 < 0.0040
Xg	=	0.0901	0.0500	<=	Xg < 0.2000
a21	=	1.211863	1.050	<=	a21 < 1.300
a31	=	0.393713	0.300	<=	a31 < 0.800

COB-1/3



**National Patent  
Analytical Systems, Inc.**

*Certificate of Calibration*

100374

Instrument Serial Number

10180

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

3-2-11

Date

Charles E. Jannard

Technician

0089/3

Final page for report

\

VANCCLSAN01\JTRAX\CASEDOCS\B-100374\B-100374\_ECF-20140

7/3/2014