

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100679

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/19/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.08: 0.080
LOT #: 11313080A1
EXPIRATION: 06/01/2015
TANK PRESSURE: 402 psi

BLANK TEST	0.000	12:35
INTERNAL STANDARD	VERIFIED	12:35
EXTERNAL STANDARD	0.079	12:35
BLANK TEST	0.000	12:36
EXTERNAL STANDARD	0.080	12:36
BLANK TEST	0.000	12:37
EXTERNAL STANDARD	0.080	12:37
BLANK TEST	0.000	12:38
EXTERNAL STANDARD	0.080	12:38
BLANK TEST	0.000	12:39
EXTERNAL STANDARD	0.080	12:39
BLANK TEST	0.000	12:40

Average = 0.0798
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.00
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 44.1°C PASSED

PUMP INFO

Flow Rate = 4.902 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

3/13/2015

Subscribed and sworn before me this 13th day of Mar. 20 15


Nikki Roth, Notary Public
My Commission Expires With Office



DMB
3/2/15