

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100353

Received From: Intoximeters

Date: 5/29/24

Analyst: Megan Crouse

Supervised by Derek Watten
JMW 05/30/24

Reason for instrument return:

Repairs for filter wheel issues.
Did not upload files since it returned from Intox.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax See 100353-2021-11-05-Change-In-Status
MNC 6/19/24

Upload N/A

Update agency information and polling status in DMHost Calling Program

Check for VOCs N/A

From Intoximeters

Intoximeters Repair Paperwork (attached)

SRO Number: SRO-057717

Received Date: 07/05/2023

Completion Date: 04/22/2024

SRO Type: REPAIR

SRO Description: AK DMT Repair

Customer ID: C000003651

Customer PO: 07052023-OBryant

Contact: Colleen OBryant

Email: Colleen.obryant@alaska.gov

Ship To:

Alaska Scientific Crime Detection Laboratory
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Units on SRO

1	LEGDMT	LEGACY DMT
---	--------	------------

Service Line: 1		**All Instruments are Calibrated to Factory Specifications**	
Unit Type:	(LEGDMT) LEGACY DMT		
Serial Number:	50100353		
Reason:	DMT Status Messages		
	FltWl er		
Reason Notes:	Customer Reported Issue: Filter wheel issue		
Resolution:	Adjusted		
	Mech adj		
Resolution Notes:	Rebuilt filterwheel assembly. Performed filter wheel tests, 100:0 pass Adjusted instrument to factory specifications.		
Parts:	<u>Qty</u>	<u>Part Number</u>	<u>Part Description</u>
	1	99-0001-00	MISC REPAIR
Reason:	DMT Functional		
	No Boot		
Reason Notes:	SBC is unserviceable. (No Comms) @ boot. SBC MAC address (xx:46:68)		
Resolution:	Replaced		
	RPL comp		
Resolution Notes:	Provided SBC from sister ticket, SRO-059515 (customer-provided) SBC MAC address (xx:33:2D)		

Order Packing Slip Report

Intoximeters
Experience • Service • Integrity

5/21/2024 7:30:03 AM

Packing Slip: 307653

Pack Date: 5/21/2024

SRO: SRO-057717



From: Intoximeters, Inc
2081 Craig Road
St Louis MO 63146
USA

Customer: C000003651

Ship To: 1

Alaska Department of Public Safety
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Alaska Scientific Crime Detection
Laboratory
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Customer PO: 07052023-OBryant

Ship Via: FDX

FED EX

Weight: 38.00

Delivery Terms: 2D

Shipping Terms: ORIGIN

Packages: 1

Customer Ship Account:

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100353

1 99-0001-00 EA 1.00 0.00
MISC REPAIR

Rec'd MNC 5/29/24