

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100675

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/26/2025

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C

PASSED

Breath Tube = 41.4°C

PASSED

PUMP INFO

Flow Rate = 4.930 L/M

PASSED

DETECTOR INFO

PUMP ON

FAILED

PUMP OFF

FAILED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

FAILED

INTERNAL STANDARD

PASSED

Diagnostic Check Failed

File only
CML 3/6/25