

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100404

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/14/2023

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.25: 0.078  
LOT #: AGL13402  
EXPIRATION: 05/14/2023  
TANK PRESSURE: 459 psi

BLANK TEST	0.000	10:50
INTERNAL STANDARD	VERIFIED	10:50
EXTERNAL STANDARD	0.078	10:50
BLANK TEST	0.000	10:51
EXTERNAL STANDARD	0.078	10:51
BLANK TEST	0.000	10:52
EXTERNAL STANDARD	0.078	10:52
BLANK TEST	0.000	10:53
EXTERNAL STANDARD	0.077	10:53
BLANK TEST	0.000	10:54
EXTERNAL STANDARD	INTERFERENCE	10:54

Interference Detected

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 47.7°C	PASSED

### PUMP INFO

Flow Rate = 4.453 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED

File only MW 05/19/23