

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Effective: 11/10/2021

Version: 2.0

Date: 1/10/22 Instrument #: 100421

Instrument Location: JBFR

Supervisor Name and Agency: C. O. Bryant / SCDL

Describe the Status Change or Issue with the Instrument:

*The instrument was brought to the lab to troubleshoot uploading. JBFR, Breath Test Supervisor, Scott Gardner, had also uploaded files from the instrument onto a USB drive + was unable to remove them from the drive. I pulled files from 12.12.2021 - 01.10.2022 from the instrument + was able to transfer them to DMHost at the lab. I also performed a non-drinking subject test + diagnostic tests (two each) + was able to transfer files to a USB drive + retrieve them as well as upload the DMT via modem line.*  
COB 1/10/22

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: ~~1/10/22~~ COB 2022.01.13

Out of Service Date/Time: 1/10/2022 COB 2022.01.13

For Use by SCDL	Additional Notes

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740