

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 11/21/2024

Instrument #: 100400

Instrument Location: Bethel Police Department

Supervisor Name and Agency: Christopher Wigner / Bethel Police Department

Describe the Status Change or Issue with the Instrument:

New instrument was sent by DPS, this unit taken out of service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time:

Out of Service Date/Time: 11/20/2024 @ 1640

For Use by SCDL

Additional Notes

100400 was taken out of service as a result of the 9/6/24 - 10/24/24 In Field Instrument Review. This instrument was being monitored for standard out and range and tank interference from the previous cycle. There were multiple status messages during the latest cycle. MNC 11/21/24

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor