

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100416

Date: 9/1/15

Analyst: Brandi Barnett

1. Reason for instrument evaluation

Instrument linearity was reading low during certification. Needs calibration.

2. Visual Inspection

3. Troubleshooting/Repair

N/A

4. VOC (attached) N/A

5. Internal Standard value $\leq 3.00\%$ Yes No N/A

6. Verify software version

Return to Factory for Repair

Ready For Calibration

Ready For Certification