

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: DMEF2014 R0
Status: Active

Instrument # 100417

Date: 4/27/15

Analyst: CRF

1. Reason for instrument evaluation

Software stuck at N.P.A.S. screen when instrument is turned on. Problem persisted after numerous restarts. -CRF 4/27/15

2. Troubleshooting/Repair

Tried loosening the screws to the screen as well as adjust connections to the screen. Unable to fix the issue. Screen was unresponsive, black or stuck on NPAS screen despite numerous restarts. CRB 5/28/15

3. VOC (attached) N/A

4. Verify software version

Return to Factory for Repair

Ready For Calibration

Ready For Certification