

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100406

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/12/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.00: 0.080  
LOT #: AG634001  
EXPIRATION: 12/05/2018  
TANK PRESSURE: 634 psi

BLANK TEST	0.000	14:20
INTERNAL STANDARD	VERIFIED	14:21
EXTERNAL STANDARD	0.079	14:21
BLANK TEST	0.000	14:22
EXTERNAL STANDARD	0.079	14:22
BLANK TEST	0.000	14:23
EXTERNAL STANDARD	0.079	14:23
BLANK TEST	0.000	14:24
EXTERNAL STANDARD	0.079	14:24
BLANK TEST	0.000	14:25
EXTERNAL STANDARD	0.079	14:25
BLANK TEST	0.000	14:26

Average = 0.0790  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C PASSED  
Breath Tube = 43.4°C PASSED

### PUMP INFO

Flow Rate = 5.071 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD


PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Charles R. Foster  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 18<sup>th</sup> day of Jan 20 19

  
Kurt R. Palmer Sr., Notary Public  
My Commission Expires With Office

  
Nikki Roth, Notary Public  
My Commission Expires With Office



DM 12/18/18