

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100418

Received From: Anchorage PD

Date: 12/23/25

Analyst: Carlisle Glaister

Reason for instrument return:

Screen issues (frozen), not corrected w/  
reboots.

### From the field

Change instrument <sup>no case over 12/23/25</sup> and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

### From Intoximeters

Intoximeters Repair Paperwork (attached) ☐

Change instrument and case location in JusticeTrax Breath Alcohol Database ☐

Reset Options and Clear Memory ☐