

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 12/19/2018 Instrument Serial Number: 100415

Instrument Location: Sitka PD

Supervisor Name and Agency: Brandi Barnett SCDL

Describe the Status Change or Issue with the Instrument:  
See attached email for documentation of Static Analysis Error on this instrument.

---

---

---

---

---

---

---

---

---

---

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: \_\_\_\_\_

-----  
**For Use by SCDL**

**Additional Notes**

---

---

---

---

---

---

---

---

---

---

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**

For questions contact the Breath Alcohol Program at 907-269-5740

## **Barnett, Brandi M (DPS)**

---

**Subject:** FW: Static Analysis Error

**From:** Barnett, Brandi M (DPS)  
**Sent:** Wednesday, December 19, 2018 9:56 AM  
**To:** 'gary.cranford@sitkapd.org'  
**Cc:** O'Bryant, Colleen S (DPS); Walton, Derek J (DPS)  
**Subject:** RE: Static Analysis Error

I looked into this and we have not seen any other static analysis errors on your instrument. The error does not have to do with static charge but it appears that it was isolated to the tests run on the one subject so we are just going to monitor your instrument. I will fill out a change in instrument status form with this information and let us know if you have any other static analysis errors or other problems.

Brandi

**From:** Foster, Charles R (DPS)  
**Sent:** Wednesday, December 19, 2018 9:20 AM  
**To:** Barnett, Brandi M (DPS)  
**Subject:** FW: Static Analysis Error

Can you look into this and respond?

**From:** Gary Cranford (Sergeant) [mailto:gary.cranford@sitkapd.org]  
**Sent:** Monday, December 17, 2018 6:53 AM  
**To:** Foster, Charles R (DPS)  
**Subject:** Static Analysis Error

Chuck,

One of my officers received the above error code yesterday (12/16) while trying to obtain a breath sample from a subject. Other than the error being referenced on page 40 of the Alaska Breath Test Operator and Supervisor Manual, I cannot find the cause of this particular code. I am thinking the Officer may have a some sort of static charge on himself and then came in contact with the instrument. Can you provide me with some feedback on this issue and if I am at least in the ballpark or not.

Sincerely,

Gary

Sergeant Gary Cranford  
Sitka Police Department  
Patrol Division  
304 Lake Street, Suite 102  
Sitka, Alaska 99835  
Office: (907) 747-3245  
Fax: (907) 747-1075