VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument State of Alaska

Serial #: 100673

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:03/08/2021

External Standard Test Values			Γ	Diagnostic Check	
EXTERNAL STANDARD INFORMATION NOMINAL: 0.080 TARGET AT 30.06: 0.080 LOT #: AG923401 EXPIRATION: 08/22/2021 TANK PRESSURE: 1029 psi			VERSIONS DMT: 3.02 PIC: 3.03 Modem: 2.6 Questions: 2.2		
BLANK TEST INTERNAL STANDARD	0.000 VERIFIED		TEMPERATURES		
EXTERNAL STANDARD	0.080		Sample Chamber = 48	B.9°C PASSED	
BLANK TEST	0.000		Breath Tube = 48		
EXTERNAL STANDARD	0.081		breach rube		
BLANK TEST	0.000		PUMP INFO		
EXTERNAL STANDARD	0.081		Flow Rate = 4.365 1	C/M PASSED	
BLANK TEST	0.000				
EXTERNAL STANDARD	0.081	09:57	DETECTOR INFO		
BLANK TEST	0.000	09:58	PUMP ON	PASSED	
EXTERNAL STANDARD	0.081	09:58	PUMP OFF	PASSED	
BLANK TEST	0.000	09:59			
			FILTER INFO		
Average = 0.0808			Filter 1	PASSED	
Std Dev = 0.0004			Filter 2	PASSED	
			Filter 3	PASSED	
			INTERNAL STANDARD	PASSED	

I, Brandi M. Barnett, after being first duly-sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.

(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program. (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Brandi M. Barnett Scientific Director State Breath Alcohol Program

Subscribed and sworn before me this

day of _____ 20 ____ Notary Public

My Commission Expires With Office





m 03/09/21