

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100348

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/24/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.65: 0.079
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 1097 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 21:44 |
| INTERNAL STANDARD | VERIFIED | 21:44 |
| EXTERNAL STANDARD | 0.074 | 21:44 |
| BLANK TEST | 0.000 | 21:45 |
| EXTERNAL STANDARD | 0.073 | 21:45 |

Standard Out of Range

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

| | |
|-------------------------|--------|
| Sample Chamber = 49.0°C | PASSED |
| Breath Tube = 45.4°C | PASSED |

PUMP INFO

| | |
|-----------------------|--------|
| Flow Rate = 4.419 L/M | PASSED |
|-----------------------|--------|

DETECTOR INFO

| | |
|----------|--------|
| PUMP ON | PASSED |
| PUMP OFF | PASSED |

FILTER INFO

| | |
|----------|--------|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

INTERNAL STANDARD

PASSED

File only DM 05/07/24