

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100413

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/01/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.59: 0.079
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 669 psi

BLANK TEST	0.000	14:37
INTERNAL STANDARD	VERIFIED	14:37
EXTERNAL STANDARD	0.077	14:37
BLANK TEST	0.000	14:38
EXTERNAL STANDARD	0.077	14:39
BLANK TEST	0.000	14:39
EXTERNAL STANDARD	0.077	14:40
BLANK TEST	0.000	14:41
EXTERNAL STANDARD	0.076	14:41
BLANK TEST	0.000	14:42
EXTERNAL STANDARD	0.077	14:42
BLANK TEST	0.000	14:43

Average = 0.0768
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
Breath Tube = 46.4°C PASSED

PUMP INFO

Flow Rate = 5.169 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Charles R. Foster
Scientific Director
State Breath Alcohol Program

9/29/17

Subscribed and sworn before me this 29 day of Sept, 2017


Kyle Klopfer, Notary Public
My Commission Expires With Office




9/19/17