

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100674

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/01/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.11: 0.081  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 1169 psi

BLANK TEST	0.000	07:55
INTERNAL STANDARD	VERIFIED	07:56
EXTERNAL STANDARD	0.080	07:56
BLANK TEST	0.000	07:57
EXTERNAL STANDARD	0.079	07:57
BLANK TEST	0.000	07:58
EXTERNAL STANDARD	0.080	07:58
BLANK TEST	0.000	07:59
EXTERNAL STANDARD	0.080	07:59
BLANK TEST	0.000	08:00
EXTERNAL STANDARD	0.079	08:00
BLANK TEST	0.000	08:01

Average = 0.0796  
Std Dev = 0.0005

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C PASSED  
Breath Tube = 46.3°C PASSED

### PUMP INFO

Flow Rate = 4.655 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

1/18/19

Charles R. Foster  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 18<sup>th</sup> day of Jan 20 19



Kurt R. Palmer Sr., Notary Public  
My Commission Expires With Office

hk 1/18/19

Nikki Roth, Notary Public  
My Commission Expires With Office

DMW 12/19/18