

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100359

Received From: Intoximeters

Date: 5/30/24

Analyst: Megan Crouse

*Supervised by Derek Watten  
JMW 05/31/24*

Reason for instrument return:

Repairs for filter wheel errors, returned from Intox.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax  N/A [See 100359-2020-08-05-Change-In-Status](#)

Upload  N/A [MNC 6/19/24](#)

Update agency information and polling status in DMHost Calling Program

Check for VOCs  N/A

From Intoximeters

Intoximeters Repair Paperwork (attached)

SRO Number: SRO-060461

Received Date: 01/17/2024

Completion Date: 05/17/2024

SRO Type: REPAIR

SRO Description: DMT Repair

Customer ID: C000003651

Customer PO: 01172024-OBryant

Contact: Colleen OBryant

Email: Colleen.obryant@alaska.gov

Ship To:

Alaska Scientific Crime Detection Laboratory  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

Units on SRO

1	LEGDMT	LEGACY DMT
---	--------	------------

Service Line: 1		<b>**All Instruments are Calibrated to Factory Specifications**</b>	
<b>Unit Type:</b>	(LEGDMT) LEGACY DMT		
<b>Serial Number:</b>	50100359		
<b>Parts:</b>	<u>Qty</u>	<u>Part Number</u>	<u>Part Description</u>
	1	59-0080-00	DMT ELEC ASSY SOLENOID FILTR WHEEL 41344
<b>Reason:</b>	DMT Status Messages		
	FltWl er		
<b>Reason Notes:</b>	Customer Reported Issue: Filter wheel error		
<b>Resolution:</b>	Replaced		
	RPL comp		
<b>Resolution Notes:</b>	Replace solenoid, run tests and everything is in spec, set to factory setting.		

# Order Packing Slip Report

**Intoximeters**  
Experience • Service • Integrity

5/21/2024 7:32:17 AM

**Packing Slip:** 307656

**Pack Date:** 5/21/2024

**SRO:** SRO-060461



**From:** Intoximeters, Inc  
2081 Craig Road  
St Louis MO 63146  
USA

**Customer:** C000003651

**Ship To:** 1

Alaska Department of Public Safety  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

Alaska Scientific Crime Detection  
Laboratory  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

**Customer PO:** 01172024-OBryant

**Ship Via:** FDX

FED EX

**Weight:** 36.00

**Delivery Terms:** 2D

**Shipping Terms:** ORIGIN

**# Packages:** 1

**Customer Ship Account:**

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100359

1	59-0080-00	EA		1.00	0.00
---	------------	----	--	------	------

DMT ELEC ASSY SOLENOID FILTR WHEEL 41344

Rec'd MNC 5/30/24