

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 9/27/17 Instrument Serial Number: 100382

Instrument Location: UAF PD

Supervisor Name and Agency: Colleen O'Bryant, SCDA

Describe the Status Change or Issue with the Instrument:

See attached email for details on RFI
status messages.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Monitored the instrument and the status message
has not occurred again. Instrument remains
in service. COB 9/27/17

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

Foster, Charles R (DPS)

From: Foster, Charles R (DPS)
Sent: Monday, June 19, 2017 3:46 PM
To: Beaudreault, Sam; 'sfgoetz@alaska.edu'
Cc: Barnett, Brandi M (DPS); O'Bryant, Colleen S (DPS)
Subject: RE: UAFPD Status message received

Good afternoon,

I just reviewed the tests from 100382 that occurred on 6/17/17. I don't think we have enough information yet to determine if this was an instrumental problem or environmental. Please periodically perform a non-drinking subject test on this Datamaster (at least once per week) and we will keep an eye on the results.

Thank you,
Chuck

From: DPS SCDL Toxicology (DPS sponsored)
Sent: Monday, June 19, 2017 8:13 AM
To: O'Bryant, Colleen S (DPS); Foster, Charles R (DPS); Barnett, Brandi M (DPS)
Subject: FW: UAFPD Status message received

From: Samuel Beaudreault [sabeaudreault@alaska.edu]
Sent: Saturday, June 17, 2017 7:43 AM
To: DPS SCDL Toxicology (DPS sponsored)
Subject: UAFPD Status message received

I received two RFI messages while testing someone this morning.
The first time, I had made sure my radio was off and there were no others in the area.
The second time, I had taken off my radio and put it out of the room about 15 feet away from the machine and it still did not work.
So I went to the Airport PD and used theirs just fine.
I just ran a test and got Interference Detected.
So now its not even the same status message happening.
What do you suggest?

I am out of the office until 6/27/17, so please respond to Chief Goetz also at sfgoetz@alaska.edu

Thanks.

--
Officer Sam Beaudreault
UAF Police Department
ph. (907)474-7721
fax (907)474-5555