



State of Alaska Scientific Crime Detection Laboratory

DataMaster Acceptance Form -DMAF2011R0

Issuing Authority - Nita Bolz, Scientific Director

Approval date: 03/01/11 | Effective date: 03/01/11 | Archive data: Active

DataMaster DMT Acceptance Form

Instrument # 100389

Date: 4/4/11

Analyst: Brandi Barnett

- 1. Check in instrument in to JusticeTrax Breath Alcohol Database
- 2. Visual Inspection
- 3. Update software
- 4. Print calibration factors (attached)
- 5. NPAS calibration paperwork (attached)
- 6. Barcode Instrument
- 7. State of Alaska Property ID# 10140884

B-100389 100389

060036100140 100389
Breath Program

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100389

03/10/2011
07:31:47

Ca = 0.1000
CAL = 0.991856 0.800 <= CAL < 1.200
b1 = 0.0003 0.0000 <= b1 < 0.0040
b2 = 0.0041 0.0010 <= b2 < 0.0100
b3 = 0.0004 0.0000 <= b3 < 0.0040
Xg = 0.0996 0.0500 <= Xg < 0.2000
a21 = 1.181601 1.050 <= a21 < 1.300
a31 = 0.429199 0.300 <= a31 < 0.800



Certificate of Calibration

100389

Instrument Serial Number

10180

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

3-14-11

Date

Charles E. Townsend

Technician

Final page for report

\

VANCCLSAN01\JTRAX\CASEDOCS\B-100389\B-100389_ECF-20140

7/3/2014