

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100345

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/16/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.47: 0.079
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 773 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 04:38 |
| INTERNAL STANDARD | VERIFIED | 04:39 |
| EXTERNAL STANDARD | 0.077 | 04:39 |
| BLANK TEST | 0.000 | 04:40 |
| EXTERNAL STANDARD | 0.077 | 04:40 |
| BLANK TEST | 0.000 | 04:41 |
| EXTERNAL STANDARD | 0.077 | 04:41 |
| BLANK TEST | 0.000 | 04:42 |
| EXTERNAL STANDARD | 0.077 | 04:42 |
| BLANK TEST | 0.000 | 04:43 |
| EXTERNAL STANDARD | 0.077 | 04:43 |
| BLANK TEST | 0.000 | 04:44 |

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 47.3°C PASSED

PUMP INFO

Flow Rate = 4.202 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

File only: Instrument being
removed from service.

BMB 12/8/17