

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Effective: 11/10/2021

Version: 2.0

Date: \_\_\_\_\_ Instrument #: \_\_\_\_\_

Instrument Location: \_\_\_\_\_

Supervisor Name and Agency: \_\_\_\_\_

Describe the Status Change or Issue with the Instrument:

---

---

---

---

---

---

---

---

---

---

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: \_\_\_\_\_

**For Use by SCDL**

Additional Notes

---

---

---

---

---

---

---

---

---

---

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**

For questions contact the Breath Alcohol Program at 907-269-5740

**O'Bryant, Colleen S (DPS)**

---

**From:** Lee, Tatiana A (DPS)  
**Sent:** Wednesday, March 9, 2022 1:40 PM  
**To:** O'Bryant, Colleen S (DPS)  
**Subject:** Fw: St Paul DataMaster VOC  
**Attachments:** SO210224100300.zip; VC210224100321.zip; VC210225100617.zip; VC210303113108.zip

---

**From:** Ann Sotelo <asotelo@stpaulak.com>  
**Sent:** Wednesday, March 2, 2022 12:02 PM  
**To:** Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov>  
**Subject:** St Paul DataMaster VOC

You don't often get email from asotelo@stpaulak.com. [Learn why this is important](#)

**CAUTION:** This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Hello Tatiana,

My name is Ann Sotelo, I am a Police Officer with Saint Paul Department of Public Safety. We recently received the new Data-Master instrument and I assisted by setting it up and conducting the VOC since, Officer Recco is out of town and in vacation. I attached the zipped files that were saved to the USB. If there's any questions, please do not hesitate to contact me.

Respectfully,

Ann Sotelo  
Police Officer  
Saint Paul Department of Public Safety  
950 Gorbach Road  
Saint Paul Island, AK 99660  
[907-600-4365](tel:907-600-4365)

**From:** [Anthony Recco](#)  
**To:** [O'Bryant, Colleen S \(DPS\)](#)  
**Subject:** Re: 100404 IS/OS  
**Date:** Tuesday, March 15, 2022 10:51:56 AM

---

I believe in service was 02/23 and out of service 03/02

Get [Outlook for iOS](#)

---

**From:** O'Bryant, Colleen S (DPS) <colleen.obryant@alaska.gov>  
**Sent:** Tuesday, March 15, 2022 1:42:40 PM  
**To:** Anthony Recco <arecco@stpaulak.com>  
**Subject:** RE: 100404 IS/OS

**I'm working to get all of this documented. Hopefully one last question. Could you tell me the dates Ann put the instrument in service and then subsequently removed it from service by unplugging it?**

---

**From:** Anthony Recco <arecco@stpaulak.com>  
**Sent:** Monday, March 14, 2022 11:56 AM  
**To:** O'Bryant, Colleen S (DPS) <colleen.obryant@alaska.gov>  
**Subject:** Re: 100404 IS/OS

She said she does not recall entering a date during set up.

Get [Outlook for iOS](#)

---

**From:** O'Bryant, Colleen S (DPS) <[colleen.obryant@alaska.gov](mailto:colleen.obryant@alaska.gov)>  
**Sent:** Monday, March 14, 2022 2:51:02 PM  
**To:** Anthony Recco <[arecco@stpaulak.com](mailto:arecco@stpaulak.com)>  
**Subject:** RE: 100404 IS/OS

**Did she change the date during set up?**

---

**From:** Anthony Recco <[arecco@stpaulak.com](mailto:arecco@stpaulak.com)>  
**Sent:** Monday, March 14, 2022 11:39 AM  
**To:** O'Bryant, Colleen S (DPS) <[colleen.obryant@alaska.gov](mailto:colleen.obryant@alaska.gov)>  
**Subject:** Re: 100404 IS/OS

**CAUTION:** This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Good afternoon,

I just talked to Ann and yes, she put the wrong year. I told her to put myself as the operator to do the set up. The Datamaster is out of service until I get back on island. Again, sorry for the confusion.

Thank you,  
Anthony Recco

Get [Outlook for iOS](#)

---

**From:** O'Bryant, Colleen S (DPS) <[colleen.obryant@alaska.gov](mailto:colleen.obryant@alaska.gov)>  
**Sent:** Monday, March 14, 2022 2:20:02 PM  
**To:** Anthony Recco <[arecco@stpaulak.com](mailto:arecco@stpaulak.com)>  
**Subject:** 100404 IS/OS

**Anthony,**

**Did you perform this tank change?**

**We have reason to believe 100404 is currently set for the incorrect year. (2021 versus 2022). Ann Sotelo emailed Tatiana indicating she put the instrument in service and attached the zipped files saved to a USB. There were 3 VOCs and 1 tank change, but you are listed as the operator.**

**I have also attached the email from Ann to Tatiana for clarification.**

**Colleen O'Bryant**

**Forensic Scientist III - Chemistry  
Alaska Scientific Crime Detection Laboratory  
4805 Dr. Martin Luther King Jr. Ave  
Anchorage, Alaska 99507**

**907-269-5783 (direct)  
907-350-2111 (cell)**

TANK CHANGE

100404 2021.02.24 Tank Change

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100404  
-----

Date: ~~02/24/2021~~\*  
Time: 10:03:00

OPERATOR NAME:  
ANTHONY T RECCO  
OPERATOR'S #: 10403

EXTERNAL STANDARD INFORMATION  
NOMINAL: 0.080  
LOT #: AG113402  
EXPIRATION: 05/14/2023

\*The date of 02/24/2021 programmed into the  
DataMaster was not correct.

Breath Test Supervisor Anthony Recco (10403)  
confirmed Breath Test Operator Ann Sotelo (9624)  
entered his name and operator number to perform this  
tank change when the instrument was placed into  
service.

COB 2022.03.14

**VERIFICATION OF CALIBRATION REPORT***of DataMaster dmt Breath Test Instrument**State of Alaska*

Serial #: 100404

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*Date: ~~02/24/2021~~\***External Standard Test Values**

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
 TARGET AT 28.97: 0.077  
 LOT #: AG113402  
 EXPIRATION: 05/14/2023  
 TANK PRESSURE: 1289 psi

BLANK TEST	0.000	10:05
INTERNAL STANDARD	VERIFIED	10:05
EXTERNAL STANDARD	0.077	10:05
BLANK TEST	0.000	10:06
EXTERNAL STANDARD	0.076	10:07
BLANK TEST	0.000	10:07
EXTERNAL STANDARD	0.076	10:08
BLANK TEST	0.000	10:09
EXTERNAL STANDARD	0.076	10:09
BLANK TEST	0.000	10:10
EXTERNAL STANDARD	0.076	10:10
BLANK TEST	0.000	10:11

Average = 0.0762  
 Std Dev = 0.0004

**Diagnostic Check**

## VERSIONS

DMT: 3.02  
 PIC: 3.02  
 Modem: 2.6  
 Questions: 2.2

## TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 48.1°C	PASSED

## PUMP INFO

Flow Rate = 4.288 L/M	PASSED
-----------------------	--------

## DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

## FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

\*Date programmed into the DataMaster was not correct.

This verification of calibration will not be signed and will not be posted online.

COB 2022.03.14

**VERIFICATION OF CALIBRATION REPORT***of DataMaster dmt Breath Test Instrument**State of Alaska*

Serial #: 100404

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*Date: ~~02/25/2021~~ \*\*

## External Standard Test Values

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
 TARGET AT 29.18: 0.078  
 LOT #: AG113402  
 EXPIRATION: 05/14/2023  
 TANK PRESSURE: 1269 psi

BLANK TEST	0.000	10:08
INTERNAL STANDARD	VERIFIED	10:08
EXTERNAL STANDARD	0.077	10:08
BLANK TEST	0.000	10:09
EXTERNAL STANDARD	0.077	10:10
BLANK TEST	0.000	10:10
EXTERNAL STANDARD	0.077	10:11
BLANK TEST	0.000	10:12
EXTERNAL STANDARD	0.077	10:12
BLANK TEST	0.000	10:13
EXTERNAL STANDARD	0.077	10:13
BLANK TEST	0.000	10:14

Average = 0.0770  
 Std Dev = 0.0000

## Diagnostic Check

## VERSIONS

DMT: 3.02  
 PIC: 3.02  
 Modem: 2.6  
 Questions: 2.2

## TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 48.1°C	PASSED

## PUMP INFO

Flow Rate = 4.395 L/M	PASSED
-----------------------	--------

## DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

## FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

\*\*Date programmed into the DataMaster was not correct.

This verification of calibration will not be signed and will not be posted online.

COB 2022.03.14

**VERIFICATION OF CALIBRATION REPORT***of DataMaster dmt Breath Test Instrument**State of Alaska*

Serial #: 100404

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*Date: ~~03/03/2021~~\*\*

## External Standard Test Values

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
 TARGET AT 29.40: 0.079  
 LOT #: AG113402  
 EXPIRATION: 05/14/2023  
 TANK PRESSURE: 1249 psi

BLANK TEST	0.000	11:33
INTERNAL STANDARD	VERIFIED	11:33
EXTERNAL STANDARD	0.077	11:33
BLANK TEST	0.000	11:34
EXTERNAL STANDARD	0.077	11:34
BLANK TEST	0.000	11:35
EXTERNAL STANDARD	0.077	11:35
BLANK TEST	0.000	11:36
EXTERNAL STANDARD	0.078	11:37
BLANK TEST	0.000	11:37
EXTERNAL STANDARD	0.077	11:38
BLANK TEST	0.000	11:39

Average = 0.0772  
 Std Dev = 0.0004

## Diagnostic Check

## VERSIONS

DMT: 3.02  
 PIC: 3.02  
 Modem: 2.6  
 Questions: 2.2

## TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 48.1°C	PASSED

## PUMP INFO

Flow Rate = 4.411 L/M	PASSED
-----------------------	--------

## DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

## FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

\*\*Date programmed into the DataMaster was not correct.

This verification of calibration will not be signed and will not be posted online.

COB 2022.03.14