

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100406

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/23/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.57: 0.079  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 1173 psi

BLANK TEST	0.000	16:05
INTERNAL STANDARD	VERIFIED	16:05
EXTERNAL STANDARD	0.079	16:05
BLANK TEST	0.000	16:06
EXTERNAL STANDARD	0.080	16:06
BLANK TEST	0.000	16:07
EXTERNAL STANDARD	0.080	16:08
BLANK TEST	0.000	16:08
EXTERNAL STANDARD	0.079	16:09
BLANK TEST	0.000	16:10
EXTERNAL STANDARD	0.079	16:10
BLANK TEST	0.000	16:11

Average = 0.0794  
Std Dev = 0.0005

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

TEMPERATURES  
Sample Chamber = 48.9°C PASSED  
Breath Tube = 43.2°C PASSED

PUMP INFO  
Flow Rate = 5.048 L/M PASSED


DETECTOR INFO  
PUMP ON PASSED  
PUMP OFF PASSED

FILTER INFO  
Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 12/7/18

Charles R. Foster  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 7 day of Dec, 2018

  
Kurt R. Palmer Sr., Notary Public  
My Commission Expires With Office



  
11/27/18