Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N ___100364___   Received From ___APD Jail___

Date ___07/12/18___   Analyst ___Derek Walton___

1. Reason for instrument return:
   Fatal error - System out of memory, rebooted itself in field.
   Undefined error on 7/7/18.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database ☑

3. Ensure change in instrument status form is in JusticeTrax ☑

4. Upload ☑

5. Update agency information and polling status in DM Host Calling Program ☑

6. Check for VOCs ☑

From Intoximeters

7. Intoximeters Repair Paperwork (attached) ☐

Approved by Forensic Alcohol Supervisor