

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100366

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/08/2016

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.94: 0.080
LOT #: 01315080A2
EXPIRATION: 02/05/2017
TANK PRESSURE: 1073 psi

BLANK TEST	0.000	04:32
INTERNAL STANDARD	VERIFIED	04:32
EXTERNAL STANDARD	0.078	04:32
BLANK TEST	0.000	04:33
EXTERNAL STANDARD	0.078	04:33
BLANK TEST	0.000	04:34
EXTERNAL STANDARD	0.078	04:34
BLANK TEST	0.000	04:35
EXTERNAL STANDARD	0.078	04:36
BLANK TEST	0.000	04:36
EXTERNAL STANDARD	0.078	04:37
BLANK TEST	0.000	04:38

Average = 0.0780
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED
Breath Tube = 42.5°C PASSED

PUMP INFO

Flow Rate = 4.804 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

File only: Instrument being removed from service due to diagnostic fail on VOCs.

BMB 6/9/16