

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100678

Received From AST/Matsu West

Date 6/8/2021

Analyst Carlisle Bailey
Derek Walton

1. Reason for instrument return:

Pump failure, screen froze, software fatal error

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload could not upload due to fatal errors

5. Update agency information and polling status in DM Host Calling Program

updated polling status & phone #, but not agency info

6. Check for VOCs unable to check for VOCs due to fatal errors

From Intoximeters

7. Intoximeters Repair Paperwork (attached)