

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100685

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/15/2019

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 43.9°C PASSED

### PUMP INFO

Flow Rate = 5.620 L/M PASSED

Communication Error 5

File only, Instrument taken out of  
service on 02/15/2019 for multiple  
status messages. JHW 03/15/2019