

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100423

Received From: Ft. WW

Date: 5/17/23

Analyst: C. O. Bryant

Reason for instrument return:

Ambient fail, gas flow error, standard out of range. Report of a subject spitting on the breath hose.

Change instrument and ~~case location~~ in JusticeTrax Breath Alcohol Database

Case present, not barcoded.

Ensure change in instrument status form is in JusticeTrax

Upload *

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)

* Attempted a VOC prior to upload (2023.05.17), it resulted in a standard out of range + was conducted at the lab. COB 5/17/23

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100423

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/17/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.95: 0.080
LOT #: 09523080A1
EXPIRATION: 05/05/2025
TANK PRESSURE: 1016 psi

BLANK TEST	0.000	11:15
INTERNAL STANDARD	VERIFIED	11:15
EXTERNAL STANDARD	0.076	11:15
BLANK TEST	0.000	11:16
EXTERNAL STANDARD	0.075	11:16
BLANK TEST	0.000	11:17
EXTERNAL STANDARD	0.074	11:17

Standard Out of Range

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 48.1°C	PASSED

PUMP INFO

Flow Rate = 4.473 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED