

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 9-29-16 Instrument Serial Number: 100373

Instrument Location: Nenana

Supervisor Name and Agency: Golden Young / AST

Describe the Status Change or Issue with the Instrument:

NO VOC DUE TO GAS FLOW ERROR

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 9-29-16 1550 hours

For Use by SCDL

Additional Notes

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dnt Breath Test Instrument*

*State of Alaska*

Serial #: 100373

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 09/29/2016

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.21: 0.081  
LOT #: AG608504  
EXPIRATION: 03/25/2018  
TANK PRESSURE: 1124 psi

BLANK TEST                    C.000    15:45  
INTERNAL STANDARD        VERIFIED    15:46

Gas Flow Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C            PASSED  
Breath Tube     = 48.1°C            PASSED

### PUMP INFO

Flow Rate = 4.756 L/M                PASSED

### DETECTOR INFO

PUMP ON                                PASSED  
PUMP OFF                                PASSED

### FILTER INFO

Filter 1                                PASSED  
Filter 2                                PASSED  
Filter 3                                PASSED

### INTERNAL STANDARD

PASSED