

Alaska Scientific Crime Detection Laboratory

DataMaster Close Out Form

Effective: 11/10/2021

Version: 2.0

Instrument #: 100350

Date: 12/28/2021

Analyst: Carlisle Glaister

Reason for Instrument Close Out

Instrument smoking after being sent to manufacturer for repair (\$ refurbishment) for the same issue.

State of Alaska Property ID#: 10140901