

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100348

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/28/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.78: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 1010 psi

BLANK TEST	0.000	17:51
INTERNAL STANDARD	VERIFIED	17:51
EXTERNAL STANDARD	0.075	17:51
BLANK TEST	0.000	17:52
EXTERNAL STANDARD	0.075	17:52
BLANK TEST	0.000	17:53
EXTERNAL STANDARD	0.075	17:54
BLANK TEST	0.000	17:54
EXTERNAL STANDARD	0.075	17:55
BLANK TEST	0.000	17:56
EXTERNAL STANDARD	0.076	17:56
BLANK TEST	0.000	17:57

Average = 0.0752
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 43.6°C	PASSED

PUMP INFO

Flow Rate = 4.315 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only DM 05/07/24