

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 7/11/2023

Instrument #: 100360

Instrument Location: 3025 Montgomery Rd, Fort Wainwright, AK 99703

Supervisor Name and Agency: LT Julayha, Sazahan Fort Wainwright Police Department

Describe the Status Change or Issue with the Instrument:

Out of Service due to multiple status messages.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time:

Out of Service Date/Time: 07/11/2023 at 1140

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Instrument pulled from service due to multiple issues: Blank Error, Ambient Fail, Static Analysis Error on Subject tests, Diagnostic Failed on VOC and filter 3 won't zero error. TLM 07/12/2023

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor