

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100348

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/19/2013

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 29.09: 0.077
 LOT #: 22612080A1
 EXPIRATION: 09/01/2014

BLANK TEST	0.000	12:24
INTERNAL STANDARD	VERIFIED	12:25
EXTERNAL STANDARD	0.078	12:25
BLANK TEST	0.000	12:26
EXTERNAL STANDARD	0.078	12:26
BLANK TEST	0.000	12:27
EXTERNAL STANDARD	0.078	12:27
BLANK TEST	0.000	12:28
EXTERNAL STANDARD	0.079	12:28
BLANK TEST	0.000	12:29
EXTERNAL STANDARD	0.078	12:29
BLANK TEST	0.000	12:30

Average = 0.0782
 Std Dev = 0.0004

Diagnostic Check

VERSIONS
 DMT: 2.00
 PIC: 3.02
 Modem: 2.3
 Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 47.6°C	PASSED

PUMP INFO

Flow Rate = 4.330 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz 12/20/13
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 20 day of Dec 20 13

Nikki Roth
 Nikki Roth, Notary Public
 My Commission Expires With Office



CRF
 12/20/13