

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100667

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/03/2020

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.12: 0.075
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 1088 psi

BLANK TEST	0.000	17:00
INTERNAL STANDARD	VERIFIED	17:01
EXTERNAL STANDARD	0.074	17:01
BLANK TEST	0.000	17:02
EXTERNAL STANDARD	0.074	17:02
BLANK TEST	0.000	17:03
EXTERNAL STANDARD	0.074	17:03
BLANK TEST	0.000	17:04
EXTERNAL STANDARD	0.074	17:04
BLANK TEST	0.000	17:05
EXTERNAL STANDARD	0.074	17:05
BLANK TEST	0.000	17:06

Average = 0.0740
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 47.5°C	PASSED

PUMP INFO

Flow Rate = 4.487 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only MR 06/04/20