

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 9/9/2023

Instrument #: 100405

Instrument Location: Fairbanks Airport Police

Supervisor Name and Agency: Vance Enderle – Fairbanks Airport Police

Describe the Status Change or Issue with the Instrument:

The instrument failed the VOC on 9/5/2023 due to a Failed Internal Standard. I have taken the DMT out of service until contacted by Breath Alcohol Staff.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 9/9/2023 at 1705 hrs

For Use by SCDL

Additional Notes

**Instrument removed from service for internal standard errors. Instrument being replaced.
COB 2023.09.11**

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor