

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 11/22/2023

Instrument #: 100687

Instrument Location: North processing room at Anchorage Jail

Supervisor Name and Agency: J. Cusack, APD, 9120

Describe the Status Change or Issue with the Instrument:

Removed from service at the request of the crime lab due to internal standard errors.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 11/22/2023 0826

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor